

**REF**

Product Code – Référence – Artikelnummer – Código – Codice

**LOT**

Lot number – Numéro de lot – Chargenbezeichnung – Lote – Numero di lotto



Use by date – A utiliser avant – Verwendbar bis – Caducidad – Data di scadenza

**STERILE EO**

Sterilized by ethylene oxide – Stérilisé à l'oxyde d'éthylène – Sterilisiert mit Ethylenoxid – Esterilizado con óxido de etileno – Sterilizzato con ossido di etilene



Do not reuse – Strict usage unique – Nur zum einmaligen Gebrauch – Válido para un solo uso – Monouso



Consult instructions for use – Lire le mode d'emploi – Lesen Sie die Gebrauchsanweisung – Leer las instrucciones de uso – Leggere le istruzioni per l'uso



Manufactured by – Fabriqué par – Hergestellt von – Fabricado por – Fabricato da



Date of manufacture – date de fabrication – Herstellungsdatum – Fecha de fabricación – Data produzione



Temperature limitation – Limite de temperature – Temperaturbegrenzung beachten – Limite de temperature – Limite di temperature



Do not use if package is damaged – Ne pas utiliser si l'emballage est endommagé – Bei beschädigter verpackung nicht verwenden – No usar en caso de envase dañado – Non utilizzare in caso di confezione danneggiata



Do not reesterilise – Ne pas restériliser – Nicht erneut sterilisieren – No reesterilizar – Non risterilizzare

**EC REP****ThermaSolutions Europe B.V.**Takkebijsters 41  
4817 BL Breda  
The Netherlands  
Phone: +31 76 579 11 44  
www.thermasolutions.com[EMEA@thermasolutions.com](mailto:EMEA@thermasolutions.com)**ThermaSolutions LLC**1889 Buerkle Road  
White Bear Lake, MN 55110  
USA  
Phone: + 1.651.209.3900  
[www.thermasolutions.com](http://www.thermasolutions.com)

info@thermasolutions.com

**CE** 1639

## Lasso and Long Lasso Drains

**Products # FLW-0010 and FLW-0111**

### Instructions for use

**Intended purpose:** The Lasso Drain and the Long Lasso Drain are intended for drainage of the abdomen. The products are referred to as "Lasso Drains" in this IFU. This IFU applies to both products.

**Mode of action:**

The function of the Lasso Drains is to drain fluids from the abdomen. The Drains are attached to tubing that will direct the fluid to reservoir or collection receptacle.

**Patient population:**

The target subject population is composed of patients undergoing lavage of the peritoneum or abdominal surgery, and the ability to tolerate general anesthesia and surgery.

**Warnings:**

1. Read the IFU in its entirety before operating the Lasso Drains. Failure to read the instructions could result in harmful effects to the user, patient, and/or Lasso Drains.
2. The Lasso Drains are STERILE. If the package is damaged, DO NOT use. Damaged packaging could compromise the sterility of the components. Replace with a new Lasso Drain and notify customer service via the contact information in this IFU.
3. Lasso Drains should only be used by a physician who has been trained in the use of the device.
4. If the Lasso Drains is combined with other equipment (such as the ThermoChem HT-2000, the user should also follow the warnings and cautions of the other device.
5. Inspect the device and sterile package carefully. Do not use if the sterile package and/or device is damaged or suspect.
6. CONFIRM expiration date on device packaging. DO NOT use if the expiration date has been exceeded.
7. The Lasso Drains is designed for SINGLE-USE only and should not be re-sterilized to avoid risk of biohazard/infection.
8. DO NOT reuse the device. If it is reused the drain may cause material damage, microbiological contamination, or biohazard/infection.
9. DO NOT modify. Product may not work as intended if altered.
10. Any medicinal substances used with the device are at the discretion of the physician. Off-label use is not promoted.

**Cautions:**

1. Ensure the Lasso Drain is connected properly and securely to connectors or tubing.
2. Store the Lasso Drains in a dry and clean environment.
3. Maintain the sterility of the components after removing from the packaging.

**Contraindications:**

1. The Lasso Drains is not indicated for use in any other areas other than the intraperitoneal area.
2. Do not use pharmaceutical or liquid solutions that are contraindicated for use or contact with PVC tubing.

**Instructions for use**

**Setup/Installation:**

1. Using sterile technique, transfer the contents of the Lasso Drain to the sterile field.
2. Select the Outflow Wye Tubing from the IPH Table Pack. The Outflow Wye Tubing is a 10" section of 3/8" tubing with a 3/8" Y connector on one end. The Outflow Connection has BLUE directional arrows.
3. Connect each end of the Lasso Tubing to Y connector. Ensure a complete insertion.
4. The Lasso Drain is ready for use.
5. Follow the instructions of the IPH Kit to complete placement of the Outflow Tubing with installed Lasso Drain Tubing to the patient.
6. Dispose of the single-use Lasso Drains per the policy of the facility.