Seacrown File No.: SC-200-104 Rev: C/0

Instruction for Use

For

Sterile Syringe for Contrast Media Injector

Compiled by	Tong qi	Date: April.,14,2018
Reviewed by	Zhong dingsu	April.,14,2018
Approved by	Zhong dingsu	April.,14,2018

Revision History Sheet

Date	Rev.	Description of change
Aug.,14,2007	A/0	Initial
Aug.,10,2011	B/0	No change for A/0
April.,14,2018	C/0	Update the new LOGO and the new typography

1. Instruction for Use

Instruction of
Sterile Syringe for Contrast Media Injector

Introduction

The Sterile Syringe for Contrast Media Injector, hereinafter called the syringe, is intended for injection of medical fluid in CT scanning, MR imaging, and angiography with the related injectors.
The medical fluids include contrast medium and saline.

Model Destination

65,100,115,150 ,200 ml

Important

Warning

-Only medical staff can handle the syringe.
-The syringe is single used. Reuse is strictly prohibited
-The syringe is used only before its valid date.
-When any broken area of the syringe package happened, it can not be used.

Caution

- Please read the instruction before use.
-The syringe for CT, and MR use should not be used in the flow rate higher than 6 ml/s. The flow rate of the injector should not be higher than 35ml/s for the syringe for DAS, and angiography use.
-Use and store the syringe at a temperature between 15℃ and 35℃.
-Store the syringe at relative humidity under 80%.

Using the Syringe

Installation

The syringe for CT, or MRI use can be installed directly in the injector head. The syringe for DSA, or angiography use should be installed with the jacket in the injector arm.
a) Make sure the plunger of injector is retrieved completely.
b) The syringe is installed on the injector to make sure a tight combination.
c) The syringe rotates 1/4 turn clockwise.

Filling

Connect one end of the fill tub to the syringe, the other to the medical fluid container. Control the injector to fill the syringe with medical fluid.

Part Connection

After the syringe is filled medical fluid, remove the J-shaped quick fill tube and connect the connector tube with the syringe and the scalp needle

Troubleshooting

a) Incapacity of handle rotation or handle rotation without contrast media
Resolution: To turn off the power source, rotate the handle of the injector contra clockwise, and then check the device and the syringe.
b) The injection must be stop when connector tube or syringe is damaged or broken, and the residual substance is removed immediately. If the device is well, a new syringe may be tried to use again.

Note: If the data of injector are set correctly, the troubles mentioned above will not happen.

Guarantee & Service

If you need service of information or if you have a problem, please visit Seacrown website at www.seacrown.cn or contact the agent of Seacrown in your country.

The syringe is guarantee for 2 years.

Understanding Symbols

Attention

Do not reuse

Sterilized with Ethylene Oxide

Use by

Production Date

Manufactured in compliance with 93/42/EEC standards

 Top

Fragile

Keep Dry

Prevent heat

European Representative

LOT number

Manufacture address

Shenzhen Seacrown Electromechnical Co., Ltd.
Add: 4F Weibaifu Building, Shekou, Nanshan District, Shenzhen, China
TEL: 86-755-26060959
FAX: 86-755-26863639
E-mail: szds@seacrown.cn
http://www. seacrown.cn
European Representative: MedNet GmbH
Address: Borkstrasse10, 48163 Münster, Germany